



Figure 1. Zero Gauss Chamber ZG12

To obtain a null field reference a Zero Gauss chamber must be used to shield the probe sensing area from external field. The ZG12 is a recommended accessory and should be ordered with all types of the SENIS Hall probes.

Dimension	[mm]	Tolerance [mm]
External diameter (D_{ext})	26	± 1.0
Internal diameter (D_{int})	21	± 0.5
Length (L)	200	± 1.0